

of our modified normal incidence silver disk pyrheliometer S. I. No. 78 against the Smithsonian APO No. 8 bis.

Three sets of data were obtained. One set was taken using the conventional procedure. The next set was obtained by using the same set of original thermometric readings as in the first case, but omitting readings c, e, and a. The third set was obtained using the amplifier-recorder equipment. Calibration of this unit was done later in the laboratory and not as would be the ideal procedure—immediately before and after the set of observations at the Smithsonian Institution. Nevertheless, results of the calibration give an agreement between the recorder results and mercury thermometer computations within about 0.1 percent.

Because use of the recorder was not contemplated in practice, no further comparisons were made.

ACKNOWLEDGMENT

The authors are indebted to Mr. L. B. Aldrich for making available technical assistance for the installation

of the thermocouple in the silver disk pyrheliometer and for supplying solar radiation flux density values for the calibration of the modified equipment.

REFERENCES

1. C. G. Abbot and F. E. Fowle, Jr., *Annals of the Astrophysical Observatory of the Smithsonian Institution*, vol. II, 1908, p. 39.
2. W. H. Hoover and A. G. Froiland, "Silver Disk Pyrheliometry," *Smithsonian Miscellaneous Collections*, vol. 122, No. 5, 1953.
3. L. B. Aldrich, "The Abbot Silver-Disk Pyrheliometer," *Smithsonian Miscellaneous Collections*, vol. 111, No. 14, 1949.
4. National Bureau of Standards, "Reference Tables for Thermocouples," *Circular 508*, May 7, 1951.
5. Ray, P. Teele, "A Measuring Circuit for Radiometer," *Journal of the Optical Society of America*, vol. 44, Nov. 1954, pp. 860-863.

CORRECTION

MONTHLY WEATHER REVIEW, vol. 83, No. 1, p. 6: Cuts for figures 9 and 10 are reversed. The one on the right should be figure 9, the one on the left, figure 10.